


Surface measurement report

Device: DTI2
 Wavelength: 653nm
 Surface quality: $\lambda/12$ ($\lambda/12$)
 Time: Sa, 30 Mär 2013 20:20:49



ID: 10mm.pdf

Error

parameter	λ	nm	status
min	-0.048	-31.293	PASSED
max	0.033	21.305	PASSED
pv	0.081	52.598	PASSED
rms	0.014	9.048	PASSED


Fit

parameter	λ	nm	status
min	-0.055	-35.797	PASSED
max	0.023	14.703	PASSED
pv	0.077	50.5	PASSED
rms	0.014	9.289	PASSED


Zernike coefficients

coefficient	value	status
0	0.06	PASSED
1	-0.006	PASSED
2	-0.115	PASSED
3	-0.02	PASSED
4	0.022	PASSED
5	-0.01	PASSED
6	-0.002	PASSED
7	-0.004	PASSED
8	-0.001	PASSED
9	0.007	PASSED
10	-0.005	PASSED
11	-0.0	PASSED
12	0.001	PASSED
13	-0.002	PASSED
14	-0.001	PASSED
15	-0.001	PASSED
16	0.001	PASSED
17	0.001	PASSED
18	-0.004	PASSED
19	0.004	PASSED
20	-0.002	PASSED
21	0.0	PASSED
22	-0.002	PASSED
23	0.0	PASSED
24	0.002	PASSED
25	-0.002	PASSED
26	-0.001	PASSED
27	-0.001	PASSED
28	0.001	PASSED
29	0.0	PASSED
30	0.0	PASSED
31	0.0	PASSED
32	0.0	PASSED
33	0.0	PASSED
34	0.0	PASSED
35	0.0	PASSED

Zernike coefficients



Surface error



Surface fit

