


Surface measurement report

Device: DTI2
 Wavelength: 653nm
 Surface quality: $\lambda/20$ ($\lambda/26$)
 Time: Sa, 30 Mär 2013 20:23:23



ID: 5mm.pdf

Error

parameter	λ	nm	status
min	-0.022	-14.138	PASSED
max	0.026	17.263	PASSED
pv	0.048	31.4	PASSED
rms	0.008	5.21	PASSED


Fit

parameter	λ	nm	status
min	-0.018	-11.911	PASSED
max	0.02	13.185	PASSED
pv	0.038	25.097	PASSED
rms	0.005	3.562	PASSED


Zernike coefficients

coefficient	value	status
0	0.047	PASSED
1	0.002	PASSED
2	-0.057	PASSED
3	-0.008	PASSED
4	0.004	PASSED
5	-0.002	PASSED
6	0.001	PASSED
7	0.0	PASSED
8	0.001	PASSED
9	0.002	PASSED
10	-0.0	PASSED
11	0.001	PASSED
12	0.0	PASSED
13	-0.002	PASSED
14	-0.001	PASSED
15	-0.008	PASSED
16	0.0	PASSED
17	-0.0	PASSED
18	-0.001	PASSED
19	-0.001	PASSED
20	-0.003	PASSED
21	0.003	PASSED
22	0.001	PASSED
23	0.001	PASSED
24	0.003	PASSED
25	0.0	PASSED
26	-0.002	PASSED
27	0.001	PASSED
28	-0.0	PASSED
29	0.0	PASSED
30	0.0	PASSED
31	0.0	PASSED
32	0.0	PASSED
33	0.0	PASSED
34	0.0	PASSED
35	0.0	PASSED

Zernike coefficients



Surface error



Surface fit

